

SEC. CL.		ORIGIN		CONTROL NO.	
C		TDS/DS		NPIC/TDS/D-927-67	
DATE OF DOC	DATE REC'D	DATE OUT	SUSPENSE DATE	CROSS REFERENCE OR POINT OF FILING	
10 July 67		139 July		997244	
TO	MEMO FOR THE RECORD				
FROM					
SUBJ.	Installation & Check-Out of the Prototype Briefing Print Enlarger				
NGA Review Complete				ROUTING	
				C/DS 11 Jul	
				DC/DS 12 Jul	
				A/TD	
				(file)	
COUNT NO.		ANSWERED		NO REPLY	
				3	

25X1

Approved For Release 2004/11/30 : CIA-RDP78B04770A000500010038-6

Approved For Release 2004/11/30 : CIA-RDP78B04770A000500010038-6

CONFIDENTIAL

957244

NPIC/TDS/D-927-67  
10 July 1967

MEMORANDUM FOR THE RECORD

SUBJECT: Installation and Check-out of the Prototype Briefing Print Enlarger

1. The prototype Briefing Print Enlarger (BPE) designed and fabricated under PAR 243, [redacted] was delivered to [redacted] on 26 June 1967. Uncrating, installation and assembly proceeded exceedingly well on that day and the equipment was under power by 1600. The balance of the week was spent in checkout to assure that the instrument performed equally as well as the in-plant check-out held at [redacted] on 20-21 June (reference Memorandum for Record, NPIC/TDS/D-905-67, 23 June 1967.)

2. Test recordings of light distribution and overall resolution of all lens sets was made. There was no case where the off-axis resolution did not meet the specification. The test target used was not retained by NPIC as it was felt that the continual use of that target through all the testing had degraded the quality. [redacted] has promised a new target set of 822 lines per millimeter and 414 lines per millimeter. The gray scale used both at [redacted] and here for check-out was retained for future checking the instrument if necessary.

3. A DD 250--Material Inspection and Receiving Report--was signed for the instrument, ten copies of the Preliminary Manual and ten sets of drawings. Seven copies of the manual as well as a like number of drawing sets will be forwarded to the recipients of the seven BPE's in fabrication under [redacted]



CONFIDENTIAL

